

1.2.4 Energy Sensors Accessories

1.2.4.1 Accessories for Pyroelectric Sensors

Fiberoptic Adapter for Pyroelectric Sensors





Oscilloscope Adapter for Pyroelectric Sensors



Heat Sink for PE-C Series Sensors



Beam Splitter Assembly




Beam splitter installed – reflected beam on sensor

Beam Splitter removed – direct beam on sensor

Beam Splitter Specifications

Material	UV grade fused silica	
Spectral range	0.19 - 2.2µm	
Aperture	Ø60mm	
Damage threshold for pulses	< 10ns PW	>300µs PW
	5J/cm ²	>200J/cm ²
Fraction split off	See graph	

F.S. Beam Splitter, 2 sided reflection unpolarized light



Accessory	Description	Part number			
Heat Sink	Heat sink that screws onto rear of PE25 and PE50 series sensors and allows working at over 50% higher average powers.	7Z08267			
Scope Adapter	Plugs in between the PE sensor and power meter. Provides BNC output to scope to see every pulse up to the maximum frequency of the sensor.	7Z11012			
Fiber Adapters	To mount fibers to sensors you need an adapter bracket and fiber adapter. All fiber adapters are compatible with the adapter bracket selected.				
Fiber Adapter Brackets	Mounting brackets to allow mounting fiber adapters to pyroelectric sensors.				
PE Sensor Family Type		Bracket P/N	Distance from fiber to detector		
PD10-C / PD10-IR-C / PD10-pj-C / PD10-IR-pj-C		7Z08275	10mm		
PE50-C / PE50BF-C		7Z08270	15mm		
PE9-C / PE9-ES-C / PE10-C / PE10BF-C / PE25-C / PE25BF-C		7Z08269	10mm		
Fiber Adapters	Fiber adapters for mounting to above brackets	SC type	ST type	FC type	SMA type
For all PE sensors above		7Z08227	7Z08226	7Z08229	1G01236A
Beam Splitter Assembly	Beam Splitter Assembly to measure pulsed laser sources too energetic for direct measurement. The reading with the Beam Splitter can be calibrated by setting the laser to a lower energy that will not damage the sensor and then taking a measurement with the beam splitter and without and taking the ratio.	7Z17001			